

CERTIFICATE OF ANALYSIS

*Leishmania infantum*

FA Substrate Slide

Catalog No.:	SLD-IFA-LSH
Lot:	P200212-001
Expiration:	17 March 2024

**Quality Control Method:** Indirect FA using *Leishmania infantum* positive control (catalog no. PC-IFA-LSH), negative control (catalog no. NC-IFA-LSH), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

**Specific Reaction:** 3-4+ signal with the positive control and negative with the negative control with no background. There are more than 50 organisms per high-power field.

**Other Comments:** NA

**QC Manager:**



David Hyatt

**Pattern Of Fluorescence:** Diffuse membranous fluorescence of organisms varying in shape from globus to trypanosome-like, some with clearly defined flagella. Single organisms and large clumps. High titer test sera will fluoresce brighter than this positive control serum.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

1 April 2020