


CERTIFICATE OF ANALYSIS

Porcine Reproductive and Respiratory Syndrome Virus European Strain (PRRSV-EU)

FA Substrate Slide

Catalog No.:	SLD-IFA-PRRS-EU
Size:	12 Well
Well Capacity:	50 µl
Lot:	PRRS-330
Expiration:	05 November 2025
Agent:	Porcine Reproductive Respiratory Syndrome Virus European Strain (PRRSV-EU)
Strain:	European Type LV
Cell Culture Substrate:	MARC cells



Description:

Wells contain virus-infected cell cultures grown on the surface of Teflon-masked slide. Slides are supplied fixed and unstained in moisture-free foil pouches. All wells contain both positive and negative cells. The positive cells usually total no more than 30% of the cells in the well so that good contrast may be seen.

Quality Control Method:

Indirect FA using PRRSV in-house European positive serum at 1/50, PRSSV negative control (catalog no. NC-IFA-PRRS), and anti-porcine IgG affinity purified FITC conjugate (catalog no. CJ-F-PORG-AP-10ML).

Specific Reaction: 3–4+ fluorescence with trace background with positive serum at 1/50. The negative control was negative with trace–1+ background. There are 0 to 15 infected cells per high-power field.

Other Comments: The monolayers in this lot have some irregularities but these have no effect on IFA performance.

Pattern Of Fluorescence:

Diffuse nuclear and cytoplasmic fluorescence with areas of bright nuclear membrane fluorescence.

Intended Use:

Generally used for Indirect FA to detect antibody to PRRSV-EU but may also be used as a positive and negative control substrate slide for Direct FA conjugates when applicable.

Storage:

Store sealed in foil pouch at -20°C. Avoid self-defrosting freezers.

References: NA

Recommended Staining Procedure for Indirect FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place diluted serum on the designated wells. Dilute serum in serum diluting buffer, pH 7.2 (catalog no. FASDB-100ML) however if high background due to anti-bovine IgG activity is present it may be advisable to use SSDB-100ML.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry around wells by pressing blotter (included in pouch) to front surface. Place labeled anti-IgG or IgM on the wells.
6. Incubate as in step 3.
7. Rinse as in step 4.
8. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
9. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be made at 400X.

Recommended Staining Procedure for Direct FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place direct FA conjugate on the designated wells.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
6. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be at 400X.

Serum Diluting Buffer (pH 7.2):*

- Na_2HPO_41.19 gm
- NaH_2PO_40.22 gm
- NaCl8.55 gm
- BSA.....10.0 gm
- DI/dH₂O.....Q.S. to 1 liter

*This recipe makes 1 liter. If you need less, adjust recipe accordingly. Store at 2-7°C. Add 0.09% NaN_3 if diluted serum is not going to be used within one week.

4X FA Rinse Buffer (pH 9.0):

- Na_2CO_311.4 gm
- NaHCO_333.6 gm
- NaCl8.5 gm
- DI/dH₂O.....Q.S. to 1 liter

Final pH should be 9.0-9.5. This is a 4X concentrate and should be diluted 1/4 with DI/distilled water for use as a working buffer. Keep in a tightly stoppered container at room temperature. MOUNTING FLUID is made by mixing glycerol and FA rinse buffer, pH 9.0, in equal proportions.

FOR *IN VITRO* LABORATORY USE ONLY.