

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P150402-007
Expiration:	21 May 2019

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-3+ fluorescence with the positive control and negative with the negative control with no background. *R. rickettsii* are too numerous to count per high-power field.

Other Comments: NA

QC Manager: *Torrey Beeson* Torrey Beeson

QC Release Date: 26 May 2015

Pattern Of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: Version # 1

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality VMRD\COAs and TDS\COA (VMRD)\8-, 10-, 12-well Slides\RMSF\SLD-IFA-RMSF_P150402-0007_190521.docx

28 May 2015