Veterinary Medical
Research & Development

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P150903-001
Expiration:	25 February 2020

Quality Control Method: IFA using R. rickettsii positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3+ signal with the positive control and negative with the negative control, no background. Structures are too numerous to count per high-power field.

Other Comments: NA

QC Manager: Will Thurman

QC Release Date: 25 February 2016

Pattern Of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

PO Box 502
Pullman, WA 99163 USA

P: 509.334.5815
F: 509.332.5356

techserve@vmrd.com
www.vmrd.com