

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P160713-001
Expiration:	22 July 2020

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-4+ signal with the positive control at neat, no background. Organisms are too numerous to count per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 25 July 2016

Pattern of Fluorescence: Fluorescent coccobacilli-like structures among ovine red blood cells. Some clumped.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\RMSF\SLD-IFA-RMSF_P160713-001_200722.docx

27 July 2016