

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P170612-001
Expiration:	21 July 2021

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-4+ signal with the positive control. Negative with trace background with the negative control. More than 25 infected cells per high powered field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 24 June 2017

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: Version #3

FOR IN VITRO LABORATORY USE ONLY.

H:\Quality\COA (VMRD)\8-, 10-, 12-well Slides\RMSF\SLD-IFA-RMSF_P170612-001_210721.docx

27 July 2017