

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P190318-001
Expiration:	30 April 2023

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 3 - 4+ signal with the positive control, trace to 1+ background. Negative with the negative control, trace to 1+ background. There are too numerous to count infected cells per high-power field.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 30 April 2019

Pattern of Fluorescence: Numerous intracellular rod-shaped bacteria.

Technical Data Sheet Version: Version # 4

FOR IN VITRO LABORATORY USE ONLY.

7 May 2019