

CERTIFICATE OF ANALYSIS

*Rickettsia rickettsii* (RMSF)


FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P190520-001
Expiration:	08 August 2023

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML).

Specific Reaction: 2-4+ signal with the positive control and negative with the negative control with trace to 1+ background. There are too numerous to count infected cells per high-power field.

Other Comments: NA

QC Manager: 

David Hyatt

QC Release Date: 08 August 2019

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: Version # 4

FOR IN VITRO LABORATORY USE ONLY.

3 September 2019