

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P190729-001
Expiration:	31 October 2023

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

Specific Reaction: 2-4+ signal with the positive control and negative with the negative control with no background. There are too numerous to count infected cells per high-power field.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: 4

FOR IN VITRO LABORATORY USE ONLY.

8 November 2019