

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P191007-001
Expiration:	04 February 2024

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-10ML).

Specific Reaction: 1-3+ signal with the positive control and negative with the negative control with trace background. Infected cells per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: 4

FOR IN VITRO LABORATORY USE ONLY.

6 February 2020