

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P220302-001
Expiration:	18 May 2026

Quality Control Method: IFA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

Specific Reaction: 3+ signal with the positive control. Negative with the negative control with trace background. Organisms per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: 4

FOR IN VITRO LABORATORY USE ONLY.

23 May 2022