

CERTIFICATE OF ANALYSIS

Rickettsia rickettsii (RMSF)

FA Substrate Slide

Catalog No.:	SLD-IFA-RMSF
Lot:	P231005
Expiration:	13 November 2031

Quality Control Method: Indirect FA using *R. rickettsii* positive control (catalog no. PC-IFA-RMSF), negative control (catalog no. NC-IFA-RMSF), and anti-canine IgG affinity purified FITC conjugate (catalog no. CJ-F-CANG-AP-10ML).

Specific Reaction: 3–4+ signal with the positive control and negative, trace background with the negative control. Organisms per high-power field are too numerous to count.

Other Comments: NA

QC Manager:



David Hyatt

Pattern Of Fluorescence: Numerous intracellular rod shaped bacteria.

Technical Data Sheet Version: 4

FOR *IN VITRO* LABORATORY USE ONLY.

17 November 2023