

CERTIFICATE OF ANALYSIS

Toxoplasma gondii


FA Substrate Slide

Catalog No.:	SLD-IFA-TOXO
Lot:	P131122-001
Expiration:	10 December 2018

Quality Control Method: Indirect FA using *Toxoplasma gondii* IgG positive control (catalog no. PC-IFA-TOXO-FEL-G), IgM positive control (catalog no. PC-IFA-TOXO-FEL-M), negative control (catalog no. NC-IFA-TOXO-FEL), anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML) and anti-feline IgM conjugate (catalog no. CJ-F-FELM-AP-1ML or 10ML).

Specific Reaction: 4+ fluorescence with the IgG positive control and 1-3+ with the IgM positive control. The diluent control was negative with no background. Organisms are too numerous to count per high powered field.

Other Comments: When using *Toxoplasma gondii* systems it is recommended to use VMRD's anti-canine affinity purified conjugate (catalog no. CJ-F-CANG-AP-1ML or 10ML) and FA Serum Diluting Buffer with 1% BSA pH 7.2 (catalog no. FASDB-100ML) as a diluent.

QC Manager:  Torrey Beeson

QC Release Date: 11 December 2014

Pattern Of Fluorescence: Organisms with bright diffuse or peripheral fluorescence.

Technical Data Sheet Version: Version # 1

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Quality VMRD\COAs and TDS\COA (VMRD)\8-, 10-, 12-well Slides\Toxoplasma\SLD-IFA-TOXO_P131122-001_181210.docx 17 December 2014