

CERTIFICATE OF ANALYSIS

Toxoplasma gondii

FA Substrate Slide

Catalog No.:	SLD- IFA-TOXO
Lot:	P190529-001
Expiration:	12 June 2023

Quality Control Method: Indirect FA using *Toxoplasma gondii* IgG positive control (catalog no. PC-IFA-TOXO-FEL-G), negative control (catalog no. NC-IFA-TOXO-FEL), anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 3-4+ signal with the positive control, no background. Negative with the negative control with apical staining. Organisms were 50 to too numerous to count per high powered field.

Other Comments: When using this slide with canine serum it is recommended to use VMRD's anti-canine affinity purified conjugate (catalog no. CJ-F-CANG-AP-1ML or 10ML) and FA Serum Diluting Buffer with 1% BSA pH 7.2 (catalog no. FASDB-100ML) as a diluent.

QC Manager:



David Hyatt

QC Release Date: 12 June 2019

Pattern Of Fluorescence: Organisms with bright diffuse or peripheral fluorescence.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

7 August 2019