

CERTIFICATE OF ANALYSIS

Toxoplasma gondii


FA Substrate Slide

Catalog No.:	SLD-IFA-TOXO
Lot:	P220415-002
Expiration:	22 June 2026

Quality Control Method: Indirect FA using *Toxoplasma gondii* IgG positive control (catalog no. PC-IFA-TOXO-FEL-G), negative control (catalog no. NC-IFA-TOXO-FEL), and anti-feline IgG affinity purified FITC conjugate (catalog no. CJ-F-FELG-AP-10ML).

Specific Reaction: 4+ signal with the positive control. Negative with apical staining with the negative control. Organisms per high-power field are too numerous to count.

Other Comments: When using *Toxoplasma gondii* systems, it is recommended to use FA Serum Diluting Buffer with 1% BSA pH 7.2 (catalog no. FASDB-100ML) as a diluent.

QC Manager: 
David Hyatt

Pattern Of Fluorescence: Organisms with bright diffuse or peripheral fluorescence.

Technical Data Sheet Version: 2

FOR *IN VITRO* LABORATORY USE ONLY.

06 July 2022