

CERTIFICATE OF ANALYSIS

Feline Infectious Peritonitis Virus Type 1 (FIP-1)

FA Substrate Slide

Catalog No.:	SLD-IFA-FIP1
Lot:	P171020-003
Expiration:	21 December 2021

Quality Control Method: Indirect FA using FIP-1 positive control (catalog no. PC-IFA-FIP1), negative control (catalog no NC-IFA-FIP1), and anti-feline IgG FITC affinity purified conjugate (catalog no. CJ-F-FELG-AP-1ML or 10ML).

Specific Reaction: 2 – 4+ signal with the positive control. Trace signal with the negative control. No background. There are 0 – 25 cells and/or syncytia per high-power field.

Other Comments: NA

QC Manager:  Will Thurman

QC Release Date: 22 December 2017

Pattern Of Fluorescence: Multiple syncytia, plaques and some individual cells with granular cytoplasmic fluorescence.

Technical Data Sheet Version: Version # 2

FOR *IN VITRO* LABORATORY USE ONLY.

H:\Master Documents\COA (VMRD)\8-, 10-, 12-well Slides\FIP-1\SLD_IFA_FIP1_P171020-003_211221.docx

08 January 2018