

CERTIFICATE OF ANALYSIS

FA Special Serum Diluting Buffer

Catalog No.:	SSDB-100ML
Lot:	P141110-002
Expiration:	19 November 2022

Quality Control Method: Indirect FA using positive controls, corresponding secondary conjugates and slides on one each of the following systems: canine CDV (catalog no. PC-IFA-CDV-G, CJ-F-CANG-1ML or 10ML, and SLD-IFA-CDV), equine EHV-1/ERV (catalog no. PC-IFA-ERV-EQ, CJ-F-EQUG-1ML or 10ML, and SLD-IFA-ERV), feline FHV-1/FVR (catalog no. PC-IFA-FVR, CJ-F-FELG-1ML or 10ML, and SLD-IFA-FVR), and porcine PCV-2 (catalog no. PC-IFA-PCV2, CJ-F-PORG-AP-1ML or 10ML, and SLD-IFA-PCV2).

Specific Reaction: This product has met VMRD, Inc. requirements. The diluent alone was negative with no background.

Other Comments: NA

QC Manager: Chris Cay



QC Release Date: 21 November 2018

Technical Data Sheet Version: Version #1

FOR *IN VITRO* LABORATORY USE ONLY.

26 November 2018