



CERTIFICATE OF ANALYSIS

Canine Parvovirus (CPV)

FA Substrate Slide

Catalog No.:	SLD-IFA-CPV
Size:	12 Well
Well Capacity:	50 µl
Lot:	P140410-002
Expiration:	04 August 2018
Agent:	Canine Parvovirus (CPV)
Strain:	Nike
Cell Culture Substrate:	CrFK TENN



Description:

Wells contain virus-infected cell cultures grown on the surface of Teflon-masked slide. Slides are supplied fixed and unstained in moisture-free foil pouches. All wells contain both positive and negative cells. The positive cells usually total no more than 30% of the cells in the well so that good contrast may be seen.

Quality Control Method:

Indirect FA using CPV IgG positive control (catalog no. PC-IFA-CPV-G), CPV in house negative control, CPV IgM positive control (catalog no. PC-IFA-CPV-M), anti-canine IgG FITC conjugate (catalog no. CJ-F-CANG-1ML or 10ML), and anti-canine IgM FITC affinity purified conjugate (catalog no. CJ-F-CANM-AP-1ML or 10ML).

Specific Reaction: 3-4+ fluorescence with the IgG positive control and 2-4+ with the IgM positive control. The diluent control was negative with no background. There are 0 to 50 infected cells per high-powered field.

Other Comments: NA

Pattern Of Fluorescence:

Single cells with cytoplasmic and nuclear fluorescence.

Interpretation Of Results:

Screen sera at 1/50 in Special Serum Diluting Buffer (SSDB-100ML) for IgG and 1/20 for IgM.

Intended Use:

Useful in detecting anti-viral antibody by IFA. May be used to differentiate antibody class (IgG or IgM) with suitable quality fluoresceinated second antibody conjugate.

Storage:

Store sealed in foil pouch at -20°C. Avoid self-defrosting freezers.

References:

Recommended Staining Procedure for Indirect FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place diluted serum on the designated wells. Dilute serum in serum diluting buffer, pH 7.2 (catalog no. FASDB-100ML) however if high background due to anti-bovine IgG activity is present it may be advisable to use SSDB-100ML.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry around wells by pressing blotter (included in pouch) to front surface. Place labeled anti-IgG or IgM on the wells.
6. Incubate as in step 3.
7. Rinse as in step 4.
8. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
9. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be made at 400X.

Recommended Staining Procedure for Direct FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place direct FA conjugate on the designated wells.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
6. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be at 400X.

Serum Diluting Buffer (pH 7.2):*

- Na₂HPO₄.....1.19 gm
- NaH₂PO₄.....0.22 gm
- NaCl.....8.55 gm
- BSA.....10.0 gm
- DI/dH₂O.....Q.S. to 1 liter

*This recipe makes 1 liter. If you need less, adjust recipe accordingly. Store at 2-7°C. Add 0.09% NaN₃ if diluted serum is not going to be used within one week.

4X FA Rinse Buffer (pH 9.0):

- Na₂CO₃.....11.4 gm
- NaHCO₃.....33.6 gm
- NaCl.....8.5 gm
- DI/dH₂O.....Q.S. to 1 liter

Final pH should be 9.0-9.5. This is a 4X concentrate and should be diluted 1/4 with DI/distilled water for use as a working buffer. Keep in a tightly stoppered container at room temperature. MOUNTING FLUID is made by mixing glycerol and FA rinse buffer, pH 9.0, in equal proportions.