

Telephone: + 1 (509) 334-5815 Fax: + 1 (509) 332-5356 E-mail: vtech@vmrd.com Web site: http://www.vmrd.com

## **Certificate of Analysis**


## CANINE DISTEMPER VIRUS (CDV)

Positive Blood Smear

**CATALOG NO.:** 214-P-CDV **LOT:** P030215001-022105

**EXPIRATION:** 21 February 2005

**VIRUS:** Canine Distemper Virus (CDV)



**DESCRIPTION:** Slide coated with blood positive for Canine Distemper Virus (CDV) antigen.

**INTENDED USE:** As a fluorescent antigen positive control training slide. Not for routine use as a control to show that the conjugate works. Rather, the CDV 2-well FA Control Slide (catalog no. 210-88-2-CDV) should be used.

**QUALITY CONTROL METHOD:** Direct FA using VMRD, Inc. CDV Direct FA Conjugate

(catalog no. 210-02-CDV).

**Specific Reaction:** 4+ positive on positive cells; negative on negative cells.

Other Reactions or Comments: Scan slides at low magnification and confirm at high magnification.

Since there are only 2-10 positive cells per slide, we recommend

testing the whole slide to ensure finding positive cells.

**PATTERN OF FLUORESCENCE:** Membrane and cytoplasmic fluorescence, including buckshot, sausage shapes, and occasional signet ring shapes in lymphocytes, polymorphonuclear cells, and monocytes.

## FOR RESEARCH AND INVESTIGATIONAL USE ONLY.

**WARRANTY:** *VMRD, Inc.* warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. *VMRD, Inc.* is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.