


CERTIFICATE OF ANALYSIS

## BOVINE ADENOVIRUS TYPE 1 (BAV-1)

Direct FA Conjugate

<b>Catalog No.:</b>	CJ-F-BAV1-10ML
<b>Volume:</b>	10 ml
<b>Lot:</b>	P121129-003
<b>Expiration:</b>	20 December 2016
<b>Agent:</b>	Bovine Adenovirus Type 1 (BAV-1)
<b>Strain:</b>	NA



### Description:

Anti-BAV-1 polyclonal antiserum conjugated to fluorescein isothiocyanate (FITC). Caprine origin. Ready to use. Liquid.

### Quality Control Method:

Direct FA using BAV-1 in house 2-well slide.

**Specific Reaction:** Conjugate produced a 2-4+ signal in the positive well. Negative on the negative well, no background.

**Other Comments:** The raw material has also been screened by Indirect FA and does not react with *Babesia bovis*, bovine adenovirus type 3 and 5 (BAV 3 and 5), bovine coronavirus (BCV), bovine leukemia virus (BLV), bovine parvovirus (BPV), bovine respiratory syncytial virus (BRSV), bluetongue virus (BTV), bovine viral diarrhea virus (BVDV), infectious bovine rhinotracheitis (IBR/BHV-1), *Neospora caninum* (bovine origin), bovine parainfluenzavirus type 3 (PI-3), bovine reovirus (REO), caprine arthritis encephalitis virus (CAEV), and *Toxoplasma gondii*.

### Pattern Of Fluorescence:

Individual cells with diffuse cytoplasmic and membrane fluorescence and strong nuclear fluorescence.

### Intended Use:

This reagent is useful for identification of BAV-1 and may react with other BAV Group A members in cell cultures and in animal tissues.

### Storage:

This conjugate is provided in liquid form and should be stored at 2-7°C. DO NOT FREEZE! It should also be stored in the original container and/or in the dark (even after dilution). If conjugate becomes cloudy it should be discarded. This conjugate contains 0.09% sodium azide as a preservative.

**References:** NA

## Recommended Staining Procedure for Indirect FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place diluted serum on the designated wells. Dilute serum in serum diluting buffer, pH 7.2 (catalog no. FASDB-100ML) however if high background due to anti-bovine IgG activity is present it may be advisable to use SSDB-100ML.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry around wells by pressing blotter (included in pouch) to front surface. Place labeled anti-IgG or IgM on the wells.
6. Incubate as in step 3.
7. Rinse as in step 4.
8. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
9. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be made at 400X.

## Recommended Staining Procedure for Direct FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place direct FA conjugate on the designated wells.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
6. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be at 400X.

## Serum Diluting Buffer (pH 7.2):\*

- Na<sub>2</sub>HPO<sub>4</sub>.....1.19 gm
- NaH<sub>2</sub>PO<sub>4</sub>.....0.22 gm
- NaCl.....8.55 gm
- BSA.....10.0 gm
- DI/dH<sub>2</sub>O.....Q.S. to 1 liter

\*This recipe makes 1 liter. If you need less, adjust recipe accordingly. Store at 2-7 C. Add 0.09% NaN<sub>3</sub> if diluted serum is not going to be used within one week.

## 4X FA Rinse Buffer (pH 9.0):

- Na<sub>2</sub>CO<sub>3</sub>.....11.4 gm
- NaHCO<sub>3</sub>.....33.6 gm
- NaCl.....8.5 gm
- DI/dH<sub>2</sub>O.....Q.S. to 1 liter

Final pH should be 9.0-9.5. This is a 4X concentrate and should be diluted 1/4 with DI/distilled water for use as a working buffer. Keep in a tightly stoppered container at room temperature. MOUNTING FLUID is made by mixing glycerol and FA rinse buffer, pH 9.0, in equal proportions.