



CERTIFICATE OF ANALYSIS

Porcine Circovirus Type 1 & 2 (PCV-1&2)

FITC Conjugate

Catalog No.:	CJ-F-PCV1&2-10ML
Volume:	10 ml
Lot:	P140530-001
Expiration:	13 June 2018
Agent:	Porcine Circovirus Type 1 & 2 (PCV-1&2)
Strain:	NA



Description:

PCV-1 and 2 polyclonal antiserum conjugated to fluorescein isothiocyanate (FITC). Porcine origin. Ready to use. Liquid.

Quality Control Method:

Direct FA using PCV-1 2-well slide (catalog no. SLD-FAC-PCV1) and PCV-2 2-well slide (catalog no. SLD-FAC-PCV2).

Specific Reaction: 3-4+ fluorescence on the positive wells of the PCV-1 and PCV-2 slides. Negative with no background on the negative wells of the PCV-1 and PCV-2 slides.

Other Comments: The raw material has also been screened by indirect FA and has been found to react with porcine adenovirus (PAV) 2-4+ with this RTU conjugate, but does not react with porcine hemagglutinating encephalomyelitis virus (PHEV), porcine parvovirus (PPV), porcine reproductive and respiratory syndrome virus (PRRSV) North American and European strains, transmissible gastroenteritis virus (TGEV), reovirus (REO), rabies Recombinant Nucleoprotein (rNP), vesicular stomatitis virus (VSV) Indiana and New Jersey strains.

Pattern Of Fluorescence:

Intense whole nuclear and variable sized but discrete cytoplasmic inclusions.

Intended Use:

Suitable for staining PCV 1&2 in cell cultures.

Storage:

This conjugate is provided in liquid form and should be stored at 2-7°C. DO NOT FREEZE! It should also be stored in the original container and/or in the dark. If conjugate becomes cloudy it should be discarded. This conjugate contains 0.09% sodium azide as a preservative.

References: NA

Recommended Staining Procedure for Indirect FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place diluted serum on the designated wells. Dilute serum in serum diluting buffer, pH 7.2 (catalog no. FASDB-100ML) however if high background due to anti-bovine IgG activity is present it may be advisable to use SSDB-100ML.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry around wells by pressing blotter (included in pouch) to front surface. Place labeled anti-IgG or IgM on the wells.
6. Incubate as in step 3.
7. Rinse as in step 4.
8. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
9. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be made at 400X.

Recommended Staining Procedure for Direct FA:

1. Warm slide to room temperature before removing from foil pouch.
2. Place direct FA conjugate on the designated wells.
3. Incubate slide in humid chamber at 37°C for 30 minutes.
4. Using a wash bottle, gently rinse slide briefly in FA rinse buffer, pH 9.0 (catalog no. FARB-4X) and then soak for 10 minutes in FA rinse buffer, pH 9.0.
5. Drain slide and dry back and edges with a paper towel. Do not allow stained surface to dry. Do not rinse with water.
6. Mount with mounting fluid [glycerol/FA rinse buffer, pH 9.0, (50/50)] (catalog no. FAMF-10ML) and view with good quality fluorescence microscope at 100X-250X. Confirmation may be at 400X.

Serum Diluting Buffer (pH 7.2):*

- Na₂HPO₄.....1.19 gm
- NaH₂PO₄.....0.22 gm
- NaCl.....8.55 gm
- BSA.....10.0 gm
- DI/dH₂O.....Q.S. to 1 liter

*This recipe makes 1 liter. If you need less, adjust recipe accordingly. Store at 2-7°C. Add 0.09% NaN₃ if diluted serum is not going to be used within one week.

4X FA Rinse Buffer (pH 9.0):

- Na₂CO₃.....11.4 gm
- NaHCO₃.....33.6 gm
- NaCl.....8.5 gm
- DI/dH₂O.....Q.S. to 1 liter

Final pH should be 9.0-9.5. This is a 4X concentrate and should be diluted 1/4 with DI/distilled water for use as a working buffer. Keep in a tightly stoppered container at room temperature. MOUNTING FLUID is made by mixing glycerol and FA rinse buffer, pH 9.0, in equal proportions.