



VMRD

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: vmrd@vmrd.com

Web site: <http://www.vmrd.com>

Certificate of Analysis

ANAPLASMA PHAGOCYTOPHILA *Negative Control Serum*

CATALOG NO.: 211-N-EE
re-order as NC-IFA-AP

VOLUME: 1 ml

LOT: P050518-002

EXPIRATION: 31 March 2012

DESCRIPTION: Equine serum diluted in PBS, 1% BSA, 0.09% sodium azide.

INTENDED USE: As negative control serum in detection of antibody to *Anaplasma phagocytophila* by indirect FA technique. This serum should be used undiluted to demonstrate non-specific fluorescence, if any.

QUALITY CONTROL METHOD: Indirect FA using *Anaplasma phagocytophila* 12 well slide (catalog no. SLD-IFA-AP), *Anaplasma phagocytophila* Positive Control (catalog no. PC-IFA-AP), Anti-Equine IgG FITC Conjugate (catalog no. CJ-F-EQUG-1ML or CJ-F-EQUG-10ML).

Specific Reaction: Negative with no background.

Other Reactions or Comments: NA

STORAGE: Store at 4°C. DO NOT FREEZE!

FOR *IN VITRO* LABORATORY USE ONLY.

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.