VMRD, Inc.
PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815
Fax: + 1 (509) 332-5356
E-mail: *vtech@vmrd.com*
Web site: *http://www.vmrd.com*

Certificate of Analysis

Ehrlichia canis

Negative Control Serum

CATALOG NO.: 211-N-EC
VOLUME: 1 ml
LOT: V02026-031205
EXPIRATION: 12 March 2005

DESCRIPTION: Canine serum diluted in PBS, 1% BSA, 0.09% sodium azide.

INTENDED USE: As negative control serum in detection of antibody to *Ehrlichia canis* by indirect FA technique. This serum should be used undiluted to demonstrate non-specific fluorescence, if any.

QUALITY CONTROL METHOD: Indirect FA using VMRD *Ehrlichia canis* Substrate Slides (catalog no. 210-88-12-EC) to detect binding; Anti-Canine IgG FITC Conjugate (catalog no. 035-10).
Specific Reaction: No specific reaction.
Other Reactions or Comments: There are some artifacts.

STORAGE: Store at 4°C. DO NOT FREEZE!

FOR RESEARCH AND INVESTIGATIONAL USE ONLY.

WARRANTY: *VMRD, Inc.* warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. *VMRD, Inc.* is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.