

**VMRD**

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: vmrd@vmrd.comWeb site: <http://www.vmrd.com>

Certificate of Analysis

Neospora caninum

*Negative Control Serum**Bovine Origin***CATALOG NO.:** 211-N-NC-BOV**VOLUME:** 1 ml**LOT:** V01125**EXPIRATION:** 15 October 2010**AGENT:** *Neospora caninum***DESCRIPTION:** Bovine sera diluted in PBS, 1% BSA, 0.09% sodium azide.**INTENDED USE:** As negative control in detection of antibody to *Neospora caninum* by indirect FA technique. This serum should be used undiluted to demonstrate non-specific fluorescence, if any.**QUALITY CONTROL METHOD:** Indirect FA using VMRD, Inc. *Neospora caninum* 12-well Substrate Slides (catalog no. SLD-IFA-NC); Anti-Bovine IgG_{1,2} FITC Conjugate (catalog no. 020-10); *Neospora caninum* Bovine Positive Control Serum (catalog no. 211-P-NC-BOV).**Specific Reaction:** Negative on negative cells.**Other Reactions or Comments:** none**PATTERN OF FLUORESCENCE:** None with our system.**STORAGE:** Store at 4-8°C. DO NOT FREEZE!

FOR <i>IN VITRO</i> LABORATORY USE ONLY.
--

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.