



**VMRD**

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: [vmrd@vmrd.com](mailto:vmrd@vmrd.com)

Web site: <http://www.vmrd.com>

## Certificate of Analysis

### ***Neospora caninum***

*Negative Control Serum*

*Canine Origin*

**CATALOG NO.:** 211-N-NC-CAN

**VOLUME:** 1 ml

**LOT:** P051213-002

**EXPIRATION:** 2 July 2011

**AGENT:** *Neospora caninum*

**DESCRIPTION:** Canine sera diluted in PBS, 1% BSA, 0.09% sodium azide.

**INTENDED USE:** As negative control in detection of antibody to *Neospora caninum* by indirect FA technique. This serum should be used undiluted to demonstrate non-specific fluorescence, if any.

**QUALITY CONTROL METHOD:** Indirect FA using VMRD, Inc. *Neospora caninum* FA Substrate Slides (catalog no. SLD-IFA-NC), *Neospora caninum* Canine Positive Control (catalog no. 211-P-NC-CAN), and Anti-Canine FITC Conjugate (catalog no. 035-1 or 035-10).

**Specific Reaction:** Negative with no background.

**Other Reactions or Comments:** NA

**PATTERN OF FLUORESCENCE:** Trace to 1+, apical staining may be visible.

**STORAGE:** Store at 4-8°C. DO NOT FREEZE!

FOR *IN VITRO* LABORATORY USE ONLY.

**WARRANTY:** VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.