



VMRD

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: vmrd@vmrd.com

Web site: <http://www.vmrd.com>

Certificate of Analysis


***Feline Viral Rhinotracheitis/
Feline Herpesvirus Type 1
(FVR/FHV-1)
Positive Control Serum***

CATALOG NO.: 211-P-FHV

VOLUME: 1 ml

LOT: P041227-002-022808

EXPIRATION: 28 February 2008



VIRUS: FVR/FHV-1

DESCRIPTION: Feline serum diluted in PBS, 1% BSA, 0.09% sodium azide.

INTENDED USE: As control serum in detection of antibody to Feline Viral Rhinotracheitis (FVR/FHV-1) by indirect FA technique. This serum should be used undiluted to demonstrate positive fluorescence.

QUALITY CONTROL METHOD: Indirect FA using VMRD, Inc. FVR/FHV-1 10-well Substrate Slides (catalog no. 210-88-10-FVR); Anti-Feline IgG FITC Conjugate (catalog no. 055-10).

Specific Reaction: 4+ at neat; endpoint is at 1:16.

Other Reactions or Comments: No background. Endpoint requires 4 further two-fold dilutions in PBS.

PATTERN OF FLUORESCENCE: Individual rounded cells with cytoplasmic and nuclear fluorescence. Some small syncytia. High titer test sera will fluoresce brighter than this positive control serum.

STORAGE: Store at 4-8°C. DO NOT FREEZE!

FOR *IN VITRO* LABORATORY USE ONLY.

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.