



VMRD

PO Box 502, Pullman, WA 99163 USA

Telephone: + 1 (509) 334-5815

Fax: + 1 (509) 332-5356

E-mail: vmrd@vmrd.com

Web site: <http://www.vmrd.com>

Certificate of Analysis


Toxoplasma gondii *Feline IgG Positive Control Serum*

CATALOG NO.: 211-P-TOXO-FEL

VOLUME: 1 ml

LOT: P060722-002

EXPIRATION DATE: 12 February 2010



DESCRIPTION: Feline serum diluted in PBS, 1% BSA, 0.09% sodium azide.

INTENDED USE: As control serum in detection of antibody to *Toxoplasma gondii* by indirect FA technique. This serum should be used undiluted to demonstrate positive fluorescence. The endpoint requires further three two-fold dilutions in PBS.

QUALITY CONTROL METHOD: Indirect FA using VMRD, Inc. *Toxoplasma* Substrate Slides (catalog no. SLD-IFA-TOXO); Anti-Feline IgG FITC Conjugate (catalog no. 055-10).

Specific Reaction: 4+ positive fluorescence at neat; endpoint is 1:16.

Other Reactions or Comments: No background

PATTERN OF FLUORESCENCE: Peripheral fluorescence of *Toxoplasma gondii* organisms. Apical (polar) staining is not considered positive. High titer test sera may fluoresce brighter than this positive control serum.

STORAGE: Store at 4-8°C.

FOR *IN VITRO* LABORATORY USE ONLY.

WARRANTY: VMRD, Inc. warrants that this product is as described in the quantity and contents stated on the label at the time of delivery to the customer. NO OTHER WARRANTIES, EXPRESS OR IMPLIED, ARE MADE BEYOND THE LABEL DESCRIPTION, INCLUDING WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR USE. Remedy is limited to replacement of the product or refund of the purchase price. VMRD, Inc. is not liable for property damage, personal injury, or economic loss caused by the product. The information listed in this information sheet is provided for reference only, and should not be substituted for the user's own incoming material quality control.